

United States Patent and Trademark Office

UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Address: COMMISSIONER FOR PATENTS P.O. Box 1450 Alexandria, Virginia 22313-1450 www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/645,739	08/21/2003	Shaher A. Ahmad	073275.0146	9156
5073	7590 09/19/2006		EXAMINER	
BAKER BOTTS L.L.P.			SWIGER III, JAMES L	
SUITE 600	2001 ROSS AVENUE SUITE 600		ART UNIT	PAPER NUMBER
DALLAS, T	X 75201-2980	3733		
·			DATE MAILED: 09/19/2006	

Please find below and/or attached an Office communication concerning this application or proceeding.

	Application No.	Applicant(s)			
	10/645,739	AHMAD, SHAHER A.			
Office Action Summary	Examiner	Art Unit			
	James L. Swiger	3733			
The MAILING DATE of this communication app	_	orrespondence address			
A SHORTENED STATUTORY PERIOD FOR REPL WHICHEVER IS LONGER, FROM THE MAILING D - Extensions of time may be available under the provisions of 37 CFR 1.1 after SIX (6) MONTHS from the mailing date of this communication If NO period for reply is specified above, the maximum statutory period - Failure to reply within the set or extended period for reply will, by statute	ATE OF THIS COMMUNICATION 36(a). In no event, however, may a reply be timwill apply and will expire SIX (6) MONTHS from	N. nely filed the mailing date of this communication.			
Any reply received by the Office later than three months after the mailin earned patent term adjustment. See 37 CFR 1.704(b).					
Status					
1)⊠ Responsive to communication(s) filed on <u>26 J</u>	une 2006.				
2a) ☐ This action is FINAL . 2b) ☑ This	This action is FINAL . 2b)⊠ This action is non-final. '				
3) Since this application is in condition for allowa	Since this application is in condition for allowance except for formal matters, prosecution as to the merits is				
closed in accordance with the practice under Ex parte Quayle, 1935 C.D. 11, 453 O.G. 213.					
Disposition of Claims					
4) ⊠ Claim(s) 1-38 is/are pending in the application 4a) Of the above claim(s) 17-29 and 33-35 is/a 5) □ Claim(s) is/are allowed. 6) ⊠ Claim(s) 1-16,30-32 and 36-38 is/are rejected. 7) □ Claim(s) is/are objected to. 8) □ Claim(s) are subject to restriction and/or	re withdrawn from consideration.				
Application Papers					
9) The specification is objected to by the Examine 10) The drawing(s) filed on 26 June 2006 is/are: a Applicant may not request that any objection to the Replacement drawing sheet(s) including the correct 11) The oath or declaration is objected to by the Examine 11.)⊠ accepted or b)□ objected to drawing(s) be held in abeyance. Sec tion is required if the drawing(s) is ob	e 37 CFR 1.85(a). jected to. See 37 CFR 1.121(d).			
Priority under 35 U.S.C. § 119					
 12) Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f). a) All b) Some * c) None of: 1. Certified copies of the priority documents have been received. 2. Certified copies of the priority documents have been received in Application No 3. Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)). * See the attached detailed Office action for a list of the certified copies not received. 					
Attachment(s)					
 Notice of References Cited (PTO-892) Notice of Draftsperson's Patent Drawing Review (PTO-948) Information Disclosure Statement(s) (PTO/SB/08) Paper No(s)/Mail Date 8/21/03;2/4/2005. 	4) Interview Summary Paper No(s)/Mail Do 5) Notice of Informal F 6) Other:	ate			

Application/Control Number: 10/645,739

Art Unit: 3733

DETAILED ACTION

Election/Restrictions

Applicant's election without traverse of invention I, claims (1-16, 30-32 and 36-38) in the reply filed on 6/26/2006 is acknowledged.

Claims 17-24, 25-29 and 33-35 are withdrawn from further consideration pursuant to 37 CFR 1.142(b) as being drawn to a nonelected invention, there being no allowable generic or linking claim. Election was made **without** traverse in the reply filed on 6/26/2006.

The species restriction has been withdrawn, directed to Figs. 4-8, in the election restriction dated 5/25/2006.

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless -

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claims 1-5, 12-13, 15-16, 30, 32, 36 and 38 are rejected under 35 U.S.C. 102(b) as being anticipated by Goble et al. (US Patent 5,417,692). Goble et al. disclose a bone anchoring device and driver comprising a screw portion (37) that penetrates bones, having a retention thread (39) on at least part of the screw portion, and a tip (40), a protrusion (44) that is adjacent the second end of the screw portion, opposite the tip, with a plurality of sides in a hexagonal shape (see Fig. 4a). Goble et al. also disclose a

Application/Control Number: 10/645,739 Page 3

Art Unit: 3733

protrusion with a hollow interior portion (19) with threads (see Fig. 10) configured to retain a component therein, and also wherein the protrusion has a maximum width that is smaller than the maximum diameter of the screw portion such that it is able to form a shoulder (see again Fig. 10). The screw portion also has a conical shape and has flutes (considered as the threads, which may cut the bone as it is inserted). They are also considered 180 degrees apart, if considered in opposing sides).

Goble et al. also disclose a driver (23) having a tip portion and a driver edge and an external surface (see Fig. 3), wherein the tip portion has a plurality of inner surfaces to match the hexagonal shape, and also wherein the bottom portion of the tip (the tapered bottom portion of 23) forms a rim outside of a second portion, which may be considered 15 (also an insert) that is capable of being inserted into the threaded bore of the bone screw portion, and wherein the rim interfaces with the shoulder of the bone anchor portion so it may be used to aid the bone anchor insertion into bone. This also has a distance that is approximately equal to the distance between the shoulder of the bone anchor and the end of the protrusion.

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negatived by the manner in which the invention was made.

Page 4

Claims 6-9 are rejected under 35 U.S.C. 103(a) as being unpatentable over Goble et al. '692. Goble et al. disclose the claimed invention except for the bone anchor shape coming in a variety of shapes. It would have been an obvious matter of design choice to one skilled in the art at the time the invention was made to construct the bone anchor having a variety of shapes to match the shape of the driver tool, since applicant has not disclosed that such a modification may solve any stated problem or is anything more than one of numerous shapes or configurations a person ordinary skill in the art would find obvious for the purpose of inserting a bone anchor using a driver tool. In re Dailey and Eilers, 149 USPQ 47 (1966).

Claim 10 is rejected under 35 U.S.C. 103(a) as being unpatentable over Goble et al. '692 in view of Kay (US Patent 5,662,683). Goble et al. disclose the claimed invention except for a bone anchor having a length that is approximately 0.20 inches. Kay disclose a bone anchor that may range from 3-18mm, which converted is 0.118 inches to 0.708 inches (see Col. 3, lines 60-67). It would have been obvious to one having ordinary skill in the art at the time the invention was made to construct the bone anchor of Goble et al. having at least the dimensions as disclosed by Kay of a range of 0.118 to 0.708 inches to better fit the anchor into bone.

Claims 11, 31 and 37 are rejected under 35 U.S.C. 103(a) as being unpatentable over Goble et al. '692 in view of McCrory et al. (US Patent 6,333,971). Goble et al. discloses the claimed invention except for a fiducial marker. McCrory et al. disclose a fiducial marker that may be anchored that allows for the device to be used with multiple imaging modalities such as CT, MRI, PET and SPECT to allow for the bone anchor to

be viewed more easily by a surgeon. It would have been obvious to one having ordinary skill in the art at the time the invention was made to construct the device of Goble et al. having at least an fiduciary marker in view of McCrory et al. to better use the device in surgery.

Claim 14 is rejected under 35 U.S.C. 103(a) as being unpatentable over Goble et al. '692. Goble et al. disclose the claimed invention except for a driver insert being made of silicon. It would have been obvious to one having ordinary skill in the art at the time the invention was made to make the insert out of silicon, since it has been held to be within the general skill of a worker in the art to select a known material on the basis of its suitability for the intended use as a matter of obvious design choice. In re Leshin, 125 USPQ 416.

Conclusion

The prior art made of record and not relied upon is considered pertinent to applicant's disclosure. See PTO-892.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to James L. Swiger whose telephone number is 571-272-5557. The examiner can normally be reached on Monday through Friday, 9:00am to 5:30pm.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Eduardo Robert can be reached on 571-272-4719. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Application/Control Number: 10/645,739

Art Unit: 3733

Page 6

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

JLS

SUPERVISORY PATENT EXAMINER